

# Certificate of Analysis

## Standard Reference Material® 1848

### Lubricating Oil Additive Package

This Standard Reference Material (SRM) is intended primarily for use in the evaluation of methods and the calibration of equipment used in the analysis of lubricating oil additive packages, engine lubricating oils, and materials of a similar matrix. SRM 1848 consists of a typical additive package used in the manufacture of crankcase lubricating oil for gasoline engines. A unit of this SRM consists of an amber, borosilicate glass bottle containing approximately 100 g of material.

**Certified Concentration Values:** The certified concentration values [1] reported as mass fractions for seven elements in SRM 1848 are listed in Table 1. A NIST certified value is a value for which NIST has the highest confidence in its accuracy in that all known or suspected sources of bias have been investigated or accounted for by NIST. The certified values are based on measurements using two or more independent analytical methods or a single NIST primary method (see Table 4).

**Reference Values:** Reference values [1] for hydrogen, nitrogen, and silicon, expressed as mass fractions, and for total base number as defined by ASTM methods D 2896 [2] and D 4739 [3] are given in Table 2. Reference values are noncertified values that are the best estimate of the true value; however, the values do not meet the NIST criteria for certification and are provided with associated uncertainties that may not include all sources of uncertainty. The reference values for total base number are method specific and apply only to the ASTM methods cited.

**Information Values:** Information values [1] are provided in Table 3 for selected physical characteristics of the material determined using selected ASTM methods. An information value is considered to be of interest to the SRM user and is provided only as additional information.

**Expiration of Certification:** The certification of this SRM lot is valid within the specified measurement uncertainties until **01 April 2010**, provided the SRM is handled and stored in accordance with the instructions given in this certificate. However, the certification will be nullified if the SRM is contaminated or modified.

**Maintenance of SRM Certification:** NIST will monitor this SRM over the period of its certification. If substantive technical changes occur that affect the certification before the expiration of this certificate, NIST will notify the purchaser. Return of the attached registration card will facilitate notification.

Coordination and overall direction of the technical measurements leading to certification of this SRM were performed by J.R. Sieber of the NIST Analytical Chemistry Division.

Statistical consultation and evaluation of measurement data were provided by S.D. Leigh of the NIST Statistical Engineering Division.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the NIST Standard Reference Materials Program by B.S. MacDonald.

Willie E. May, Chief Analytical Chemistry Division

Gaithersburg, MD 20899

Nancy M. Trahey, Chief
Certificate Issue Date: 20 September 2000

Standard Reference Materials Program
NIST analytical measurements were performed by W.R. Kelly, E.A. Mackey, J.L. Mann, A.F. Marlow, J.R.

SRM 1848 Page 1 of 4

Sieber, R.D. Vocke, Jr., and L.J. Wood of the NIST Analytical Chemistry Division.

#### **Collaborating Laboratories**

The following laboratories contributed to the certification and physical property characterization of this SRM.

BP Amoco, Naperville, IL Chevron Research and Technology Company, Richmond, CA Exxon Research and Engineering Company, Annandale, NJ Phillips Petroleum Company, Bartlesville, OK Texaco Technology Ghent, Ghent, Belgium The Lubrizol Corporation, Wickliffe, OH

**Source of Material:** The material used for this SRM was prepared by a commercial supplier of lubricant additives according to its normal procedures and specifications. The material was bottled at NIST under the supervision of D.G. Friend of the NIST Standard Reference Materials Program.

**Use and Storage:** To relate analytical determinations to the certified values on this Certificate of Analysis, a minimum sample mass of 30 mg is recommended. The material should be stored in its original container at room temperature. If the material has not been sampled recently, it may be advisable to stir the contents of the bottle using a clean glass or plastic implement.

Table 1. Certified Concentrations (Mass Fractions, in %) for Selected Elements

Concentr	ation
0.137 ±	0.019
0.359 ±	0.011
$0.0927~\pm$	0.0020
0.821 ±	0.038
$0.788 \pm$	0.028
$2.3270~\pm$	0.0043
0.873 ±	0.022
	Concentra 0.137 ± 0.359 ± 0.0927 ± 0.821 ± 0.788 ± 2.3270 ± 0.873 ±

The certified value for each element, except sulfur, is equal to a weighted mean of the results of the cited methods performed by NIST and industry laboratories. The uncertainty listed with each value is an expanded uncertainty based on a 95 % confidence interval [4] calculated as  $U = ku_c$ , where  $u_c$  is the combined standard uncertainty and k = 2 is a coverage factor. The expanded uncertainty is calculated by combining a between-method variance [5] with a pooled, within-method variance following the ISO Guide [6]. For sulfur, the certified value is based on a single method, with the expanded uncertainty including an estimate of the bias in the method.

Table 2. Reference Values for Selected Elements and Total Base Number

Element or Property	Con	centration
Hydrogen	12.3 %	± 0.4 % (mass fraction)
Nitrogen	0.57 %	$\pm$ 0.03 % (mass fraction)
Silicon	0.0050 %	± 0.0002 % (mass fraction)
Total Base Number by ASTM D 2896*	56.7	$\pm 0.7 \text{ mg KOH/g}$
Total Base Number by ASTM D 4739*	49.6	$\pm$ 5.6 mg KOH/g

<sup>\*</sup>This property is method specific. The value stated applies only to the ASTM method cited using the appropriate quantity of material specified for that test.

Each reference value is equal to a weighted mean of the results of the cited methods performed by NIST or industry

SRM 1848 Page 2 of 4

laboratories. The uncertainty listed with each value is an expanded uncertainty [4] based on a 95 % confidence interval calculated as  $U = ku_c$ , where  $u_c$  is the combined standard uncertainty and k = 2 is a coverage factor. The expanded uncertainty is calculated by combining a between-method variance [5] with a pooled, within-method variance following the ISO Guide [6]. Reference values for hydrogen and silicon are the mean from a single method for each element. For those elements, the expanded uncertainty includes an estimate of the bias in the method.

Table 3. Information Values for SRM 1848

Property	ASTM Method Used	Value
COC Flash Point	D 92 [7]	194 °C
Pensky-Martens Flash Point	D 93 [8]	173 °C
Kinematic Viscosity, 100 °C	D 445 [9]	170 cSt
Kinematic Viscosity, 40 °C	D 445 [9]	4000 cSt
Ash	D 482 [10]	5.0 % (mass fraction)
Total Acid Number	D 664 [11]	21 mg KOH/g
Sulfated Ash	D 874 [12]	6.1 % (mass fraction)

Table 4. Methods of Analysis for SRM 1848 (Certified Elements Are in Bold)

Element	Methods Used
Boron	PGAA [13], ICPAES, ASTM D 4951 [14], ASTM D 5185 [15]
Calcium	ICPAES, ASTM D 4927 [16]
Chlorine	XRF, PGAA, DIN 51577 [17]
Hydrogen	PGAA
Magnesium	XRF, ICPAES, ASTM D 4951
Nitrogen	ASTM D 5291 [18], ASTM D 5762 [19]
Phosphorus	XRF, ICPAES, ASTM D 4047 [20], ASTM D 4927, Compleximetric tration
Silicon	XRF
Sulfur	ID-TIMS [21]
Zinc	PGAA, ICPAES, ASTM D 4951, ASTM D 4927

#### **NIST Methods**

PGAA	Prompt gamma-ray activation analysis
<b>ICPAES</b>	Inductively coupled plasma atomic emission spectrometry
<b>ID-TIMS</b>	Isotope dilution - thermal ionization mass spectrometry
XRF	X-ray fluorescence spectrometry: DIN 51577 Part 2, ASTM D 4927 Method A, and other internal
	standard procedures

SRM 1848 Page 3 of 4

#### REFERENCES

- [1] May, W.E., Gills, T., Parris, R., Beck, II, C., Fassett, J.D., Gettings, R., Greenberg, R.D., Guenther, F., Kramer, G., MacDonald, B.S., and Wise, S., "Definitions of Terms and Modes Used at NIST for Value-Assignment of Reference Materials for Chemical Measurements," NIST Special Publication 260-136, (1999).
- [2] ASTM D 2896, Annual Book of ASTM Standards, Vol. 05.02, West Conshohocken, PA, (1999).
- [3] ASTM D 4739, Annual Book of ASTM Standards, Vol. 05.02, West Conshohocken, PA, (1999).
- [4] Hahn, G.J. and Meeker, W.Q. "Statistical Intervals: A Guide for Practitioners," John Wiley & Sons, Inc., New York, (1991).
- [5] Rukhin, A.L. and Vangel, M.G., "Estimation of a Common Mean and Weighted Means Statistics," Journal of the American Statistical Association, **Vol. 93** (441), p. 303 (1998).
- [6] Guide to the Expression of Uncertainty in Measurement, ISBN 92-67-10188-9, 1st Ed. ISO, Geneva, Switzerland, (1993); see also Taylor, B.N. and Kuyatt, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, U.S. Government Printing Office, Washington DC, (1994); available at <a href="http://physics.nist.gov/Pubs/">http://physics.nist.gov/Pubs/</a>.
- [7] ASTM D 92-97, Annual Book of ASTM Standards, Vol. 05.01, West Conshohocken, PA, (1999).
- [8] ASTM D 93-97, Annual Book of ASTM Standards, Vol. 05.01, West Conshohocken, PA, (1999).
- [9] ASTM D 445-97, Annual Book of ASTM Standards, Vol. 05.01, West Conshohocken, PA, (1999).
- [10] ASTM D 482-95, Annual Book of ASTM Standards, Vol. 05.01, West Conshohocken, PA, (1999).
- [11] ASTM D 664-95, Annual Book of ASTM Standards, Vol. 05.01, West Conshohocken, PA, (1999).
- [12] ASTM D 874-96, Annual Book of ASTM Standards, Vol. 05.01, West Conshohocken, PA, (1999).
- [13] Mackey, E.A., Gordon, G.E., Lindstrom, R.M., and Anderson, D.L., "Effects of Target Shape and Neutron Scattering on Element Sensitivities for Neutron Capture Prompt Gamma-ray Activation Analysis," Analytical Chemistry Vol. 63, pp. 288-292, (1991).
- [14] ASTM D 4951-96, Annual Book of ASTM Standards, Vol. 05.03, West Conshohocken, PA, (1999).
- [15] ASTM D 5185-97, Annual Book of ASTM Standards, Vol. 05.03, West Conshohocken, PA, (1999).
- [16] ASTM D 4927-96, Annual Book of ASTM Standards, Vol. 05.02, West Conshohocken, PA, (1999).
- [17] DIN 51577, Part 2, English, Deutsches Institut Fur Normung, Beuth Verlag GmbH, Berlin, (1993).
- [18] ASTM D 5291-96, Annual Book of ASTM Standards, Vol. 05.03, West Conshohocken, PA, (1999).
- [19] ASTM D 5762-98, Annual Book of ASTM Standards, Vol. 05.03, West Conshohocken, PA, (1999).
- [20] ASTM D 4047-91, Annual Book of ASTM Standards, Vol. 05.02, West Conshohocken, PA, (1999).
- [21] Kelly, W.R., Paulsen, P.J., Murphy, K.E., Vocke, R.D. Jr., and Chen, L.-T., "Determination of Sulfur in Fossil Fuels by Isotope Dilution Thermal Ionization Mass Spectrometry," Analytical Chemistry **Vol. 66**, p. 2505 (1994).

Users of this SRM should ensure that the certificate in their possession is current. This can be accomplished by contacting the SRM Program at: telephone (301) 975-6776; fax (301) 926-4751; e-mail srminfo@nist.gov; or via the Internet <a href="http://www.nist.gov/srm">http://www.nist.gov/srm</a>.

SRM 1848 Page 4 of 4